



August 8, 2022

## **CPSC Urges Consumers to Immediately Stop Use of Mobility Transfer Systems Adult Portable Bed Rails Due to Entrapment and Asphyxia Hazard**

The U.S. Consumer Product Safety Commission (CPSC) is warning consumers to immediately stop using 285,000 Mobility Transfer Systems adult portable bed rails. According to CPSC, these products can create an entrapment hazard and pose a risk of serious injury or death to users. CPSC reports that the bed rails were manufactured and sold by Mobility Transfer Systems Inc. from 1992 to 2021, and by Metal Tubing USA Inc. in 2021 and 2022.

The warning includes the following products:

Freedom Grip Economy Bed Handle, Freedom Grip Transfer Bedrail, Reversible Slant Rail, Freedom Grip Plus Bed Rail, Support Bar Assist Handle, MTS FREEDOM Grip Bed Rail/Bed Handle, MTS Medical Supply Mobility Transfer Systems Transfer Handle, 15 Pound, Freedom Grip® Plus Bed Handle, Complete Medical MTS600 Reversible Slant Rail for Home Beds, FREEDOM Grip Bed Rail, Support Bar Assist Handle, Security Rails - Electric Style Beds - Double Sided - 18" Length.

**Consumers should immediately stop using the items.**

**More details about these and other products can be found in the following notification:** <https://www.cpsc.gov/Newsroom/News-Releases/2022/CPSC-Urges-Consumers-to-Immediately-Stop-Use-of-Mobility-Transfer-Systems-Adult-Portable-Bed-Rails-Due-to-Entrapment-and-Asphyxia-Hazard-Three-Deaths-Reported>.

If you have any questions regarding this notice, you may contact Maryland Access Point, Howard County at [410-313-1234](tel:410-313-1234) (voice/relay) or by email at [map@howardcountymd.gov](mailto:map@howardcountymd.gov).